



HAL
open science

3D Visualization of intraoperative stimulation test results for better target selection in DBS surgery

Ashesh Shah, Fabiola Alonso, Jean-Jacques Lemaire, Karin K Wårdell, Daniela Pison, Jerome Coste, Erik Schkommodau, Simone Hemm-Ode

► To cite this version:

Ashesh Shah, Fabiola Alonso, Jean-Jacques Lemaire, Karin K Wårdell, Daniela Pison, et al.. 3D Visualization of intraoperative stimulation test results for better target selection in DBS surgery. Meeting of the Swiss, Austrian and German Societies of Biomedical Engineering, Swiss, Austrian and German Societies of Biomedical Engineering, Oct 2016, Basel, Switzerland. pp.S243. hal-01658519

HAL Id: hal-01658519

<https://uca.hal.science/hal-01658519>

Submitted on 8 Nov 2018

HAL is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.

3D Visualization of intraoperative stimulation test results for better target selection in DBS surgery

Shah A., Alonso F., Pison D., Lemaire J.-J., Wårdell K., Coste J., Schkommodau E., Hemm-Ode S.

Ashesh Shah, Institute for Medical and Analytical Technologies, University of Applied Sciences and Art Northwestern Switzerland, Muttenz, Basel, ashesh.shah@fhnw.ch

Fabiola Alonso, Department of Biomedical Engineering, Linköping University, Linköping, Sweden, fabiola.alonso@liu.se

Daniela Pison, Institute for Medical and Analytical Technologies, University of Applied Sciences and Art Northwestern Switzerland, Muttenz, Basel, Daniela.pison@fhnw.ch

Jean-Jacques Lemaire, Centre Hospitalier Universitaire de Clermont-Ferrand, Image-Guided Clinical Neurosciences and Connectomics, Université d'Auvergne, France, jjlemaire@chu-clermontferrand.fr

Karin Wårdell, Department of Biomedical Engineering, Linköping University, Linköping, Sweden, karin.wardell@liu.se

Jérôme Coste, Centre Hospitalier Universitaire de Clermont-Ferrand, Image-Guided Clinical Neurosciences and Connectomics, Université d'Auvergne, France, jcoste@chu-clermontferrand.fr

Erik Schkommodau, Institute for Medical and Analytical Technologies, University of Applied Sciences and Art Northwestern Switzerland, Muttenz, Basel, erik.schkommodau@fhnw.ch

Simone Hemm-Ode, Institute for Medical and Analytical Technologies, University of Applied Sciences and Art Northwestern Switzerland, Muttenz, Basel; Department of Biomedical Engineering, Linköping University, Linköping, Sweden, simone.hemm@fhnw.ch

Despite an increasing use of deep brain stimulation (DBS), the fundamental mechanisms underlying therapeutic and adverse effects as well as the optimal stimulation site remain largely unknown. The aim of the present study was to develop a method to correlate electric field simulations for intraoperative stimulation tests with quantitatively evaluated symptom improvement and patient specific anatomy to estimate the optimal stimulation site. One essential tremor patient, bilaterally implanted in the ventro-intermediate nucleus (Vim) has been included. Preoperatively Vim and its anatomic neighbors were manually outlined according to spontaneous MRI contrasts using the commercial planning software from where the structures could be exported via a specifically designed interface. During the intervention, intraoperative stimulation tests were performed on two trajectories per hemisphere (8 positions per trajectory). The change in tremor at each stimulation position compared to baseline was evaluated using a 3-axis accelerometer. Based on accelerometer data, two stimulation amplitudes (low and high improvement) were identified per position and corresponding electric-field isosurfaces (0.2V/mm) were simulated. As each voxel in the region of interest may be part of several isosurfaces -each surface depicting one amplitude responsible for one improvement in tremor- the voxel was assigned to the isosurface representing the minimum improvement using Matlab. Data were imported into Paraview (VTK based 3D visualization software). Color-coded minimum 3D-improvement maps were visualized on the patient's MR images together with the manually outlined anatomical structures. The resulting visualization was evaluated by clinicians. The software allowed 3D visualization as well as orthographic slices parallel to the trajectory. Clinicians confirmed that it enables the identification of the most effective stimulation areas with respect to the anatomy. This new concept based on quantitative symptom evaluation, electric field simulations, and patient specific anatomical data will allow the analysis of a high amount of intraoperative data which might help to elucidate the mechanism of action of DBS.