



HAL
open science

Co-registering Thalamus for Analysis of Patient-specific Intra-operative Improvement Maps in DBS

Dorian Vogel, Ashesh Shah, Fabiola Alonso, Jean-Jacques Lemaire, Jerome Coste, Karin Wårdell, Simone Hemm

► To cite this version:

Dorian Vogel, Ashesh Shah, Fabiola Alonso, Jean-Jacques Lemaire, Jerome Coste, et al.. Co-registering Thalamus for Analysis of Patient-specific Intra-operative Improvement Maps in DBS. 23th Congress of the European Society for Stereotactic and Functional Neurosurgery, European Society for Stereotactic and Functional Neurosurgery (ESSFN), Sep 2018, Edimbourg, United Kingdom. hal-01883401

HAL Id: hal-01883401

<https://uca.hal.science/hal-01883401v1>

Submitted on 28 Sep 2018

HAL is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.

Co-registering Thalamus for Analysis of Patient-specific Intra-operative Improvement Maps in DBS

ESSFN 2018 | Poster P094 Movement Disorders

Dorian VOGEL (Speaker), Ashesh SHAH, Fabiola ALONSO, Jean-Jacques LEMAIRE, Jérôme COSTE, Karin WÅRDELL, Simone HEMM

Deep brain stimulation (DBS) is used for symptomatic treatment of movement disorders. In many centers, the optimal DBS position is identified by intra-operative stimulation tests, which are performed at several positions along the planned trajectories toward the target structure with various stimulation amplitudes. The aim of the study was to propose a methodology summarizing all available intra-operative data of different patients to an atlas to provide targeting guidelines in the thalamus.

Data was collected for 8 patients (6 Essential Tremor, 2 Parkinson's Disease) who underwent bilateral implantation toward the ventral intermediate nucleus (Vim) (16 trajectories). Stereotactic MRI (T1, WAIR [1]; 1.5T, Siemens Sonata) were acquired prior to DBS surgery. An acceleration sensor was mounted on the contralateral wrist of the patient during intraoperative test stimulation to quantify tremor improvement. Written informed consent was obtained from the patients (ref.: 2011-A00774-37/AU905). A conductivity model of each patient's brain was derived from the intensity values of the T1 image. Finite element simulations were performed using these models for the positions and amplitudes (between 0 and 3 mA) with the best improvement. A 0.2 V/mm isolevel [2] was used to visualize the electric field. Improvement maps of each patient's brain were visualized by affecting to each voxel in simulation results the maximum improvement measured. These maps were further enhanced by 14 subthalamic structures per side labeled by the neurosurgeon [3]. The resulting Improvement Map was visualized together with the patients' anatomy and the implant position providing a summary of the intraoperative testing [4]. An example view of this visualization is presented in Figure 1. In order to proceed to group analysis, a normalization pipeline for the improvement maps was designed as a two-step process: the first step uses the MNI152 template [5] and tools from FSL [6] to proceed to an initial linear registration. The T1 image for each patient was aligned and then registered to the MNI template. The WAIR image was then aligned to the aligned T1 and the registration transformation was applied. Both transforms were also applied to the anatomical structures. A result of this first normalization is presented in Figure 2 for the ventro intermediate nucleus (Vim). Those results show the necessity of a finer registration step, allowing local displacement of voxels in order to provide a better definition of the anatomical labels after normalization. In the second step, several normalization templates as well as non-linear group registration tools were evaluated in order to obtain the best probabilistic definition of the labeled sub-thalamic structures after application of the disformation field. Future work will focus on transferring the improvement values and spatial distribution of the electric field to the anatomical atlas to create an improvement atlas, and select appropriate statistical methods to visualize it. The atlas will present a summary of the complete patient pool, to be used as an assistance tool during planning. As the number of patient increases, this type of atlas will help analyzing the mechanisms of action.

[1] F. Vassal *et al.*, "Direct stereotactic targeting of the ventrointermediate nucleus of the thalamus based on anatomic 1.5-T MRI mapping with a white matter attenuated inversion recovery (WAIR) sequence," *Brain Stimulat.*, Oct. 2012.

[2] M. Åström, E. Diczfalusy, H. Martens, and K. Wårdell, "Relationship between Neural Activation and Electric Field Distribution during Deep Brain Stimulation," *IEEE Trans. Biomed. Eng.*, Feb. 2015.

[3] J.-J. Lemaire, L. Sakka, L. Ouchchane, F. Caire, J. Gabrillargues, and J.-M. Bonny, "Anatomy of the Human Thalamus Based on Spontaneous Contrast and Microscopic Voxels in High-Field Magnetic Resonance Imaging," *Oper. Neurosurg.*, Mar. 2010.

[4] A. Shah *et al.*, *Improving DBS targeting using 3D visualization of intraoperative stimulation tests*, vol. 94. European Society for Stereotactic and Functional Neurosurgery (ESSFN), 2016.

[5] G. Grabner, A. L. Janke, M. M. Budge, D. Smith, J. Pruessner, and D. L. Collins, "Symmetric Atlasing and Model Based Segmentation: An Application to the Hippocampus in Older Adults," in *Medical Image Computing and Computer-Assisted Intervention*, 2006.

[6] S. M. Smith *et al.*, "Advances in functional and structural MR image analysis and implementation as FSL," *NeuroImage*, Jan. 2004.