



**HAL**  
open science

## Effect of water temperature on a phycocyanin probe for monitoring cyanobacteria and its effectiveness on biomass estimation of three cyanobacteria morphotypes

Laurianne Paris, Mathilde Dutaut, Sylvain Coly, David Foltier, Benoît Gire, Marianne Coulon, Antoine Thouvenot, Delphine Latour, Benjamin Legrand

### ► To cite this version:

Laurianne Paris, Mathilde Dutaut, Sylvain Coly, David Foltier, Benoît Gire, et al.. Effect of water temperature on a phycocyanin probe for monitoring cyanobacteria and its effectiveness on biomass estimation of three cyanobacteria morphotypes. Phycotox 2023 : Conférence annuelle GdR Phycotox - GIS Cyanobactéries, May 2023, Nantes, France. hal-04282414

**HAL Id: hal-04282414**

**<https://uca.hal.science/hal-04282414>**

Submitted on 13 Nov 2023

**HAL** is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.



Distributed under a Creative Commons Attribution 4.0 International License

## Effect of water temperature on a phycocyanin probe for monitoring cyanobacteria and its effectiveness on biomass estimation of three cyanobacteria morphotypes

Laurianne Paris<sup>1</sup>, Mathilde Dutaut<sup>2</sup>, Sylvain Coly<sup>2,3</sup>, David Foltier<sup>2</sup>, Benoît Gire<sup>2</sup>, Marianne Coulon<sup>1,2</sup>, Antoine Thouvenot<sup>2</sup>, Delphine Latour<sup>1</sup>, Benjamin Legrand<sup>2</sup>

<sup>1</sup> Université Clermont Auvergne, CNRS, LMGE, F-63000 Clermont-Ferrand, France

<sup>2</sup> ATHOS Environnement, 112 Avenue du Brézet, 63100 Clermont Ferrand, France

<sup>3</sup> Université Clermont Auvergne, CNRS, LMBP, F-63000 Clermont-Ferrand, France

E-mail: [laurianne.paris@uca.fr](mailto:laurianne.paris@uca.fr)

---

In a context of global warming and high anthropization of freshwater systems, cyanobacteria monitoring becomes a key priority. Thus, water companies have recently deployed large-scale submersible fluorescence sensors, capable of measuring diagnostic pigments and estimating cyanobacteria biomass in near-real time. However, these tools have several limitations, such as the detection limit of phycocyanin or the impact of temperature, one of the most important factors generating a bias on the measurement of phycocyanin probes. In an attempt to address these shortcomings, we compared the efficiency of a phycocyanin probe at different temperatures, first on three cyanobacteria cultures (*Microcystis*, *Dolichospermum* and *Synechocystis*) and secondly on a small eutrophic lake in continuous recording.

The *in vitro* results showed a high correlation between cyanobacterial biomasses, estimated by flow cytometry (*Microcystis* and *Synechocystis*) or by microscopic counting (*Dolichospermum*), and phycocyanin probe measurement, with  $R^2$  higher than 0.96, even for biomasses lower than hundreds of  $\text{mm}^3 \cdot \text{L}^{-1}$ . A saturation effect was demonstrated only between 350 and 550  $\text{mm}^3 \cdot \text{L}^{-1}$  depending on the strains, corresponding to very high concentrations, far from the limit value for the use of water. However, laboratory tests on temperature highlighted the limitations of the sensor with a strong overestimation and underestimation at low ( $< 15 \text{ }^\circ\text{C}$ ) and high ( $> 25 \text{ }^\circ\text{C}$ ) temperatures, respectively. Given this discrepancy, a temperature correction factor was determined and applied over the three years of field data, which significantly improved the signal quality. The strong relationship found between the corrected signal and the cyanobacteria biovolume obtained by microscopic enumeration allowed us to validate this temperature correction factor, offering the possibility of using it for monitoring other lakes and to better understand the short-term dynamics of cyanobacteria.

**Keywords:** Cyanobacteria; Probe sensor monitoring; Phycocyanin; Temperature; Picocyanobacteria