



Characterizing Critical Buried Interfaces by Inelastic Background Analysis (IBA) using Lab-Scale Hard X-ray Spectroscopy

Taylor Rose Bure, Christine Robert-Goumet, Olivier Renault, Emmanuel Nolot, Thierry Conard, Nicolas Pauly

► To cite this version:

Taylor Rose Bure, Christine Robert-Goumet, Olivier Renault, Emmanuel Nolot, Thierry Conard, et al.. Characterizing Critical Buried Interfaces by Inelastic Background Analysis (IBA) using Lab-Scale Hard X-ray Spectroscopy. LES JOURNEES NATIONALES DES SPECTROSCOPIES DE PHOTOEMISSION (JNSPE) 2024, May 2024, Lille, France. hal-04628141

HAL Id: hal-04628141

<https://uca.hal.science/hal-04628141v1>

Submitted on 28 Jun 2024

HAL is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.



Distributed under a Creative Commons Attribution 4.0 International License

Characterizing Critical Buried Interfaces by Inelastic Background Analysis (IBA) using Lab-Scale Hard X-ray Spectroscopy

T. R. Bure¹, C. Robert-Goumet¹, O. Renault², E. Nolot², T. Conard³, N. Pauly⁴

¹Université Clermont Auvergne, Clermont Auvergne INP, CNRS, Institut Pascal, F-63000 Clermont-Ferrand, France

²Univ. Grenoble Alpes, CEA, Leti, F-38000 Grenoble, France

³IMEC, Kapeldreef 75, B-3001, Leuven, Belgium


⁴Université Libre de Bruxelles, Service de Métrologie Nucléaire (CP 165/84), 50 av. F. D. Roosevelt, B-1050 Brussels, Belgium




UNIVERSITÉ
LIBRE
DE BRUXELLES

imec

ABSTRACT: devices are shrinking while increasing in complexity. The buried interfaces in electronic materials are critical to their functionality but notoriously difficult to characterize. This work aims to expand an existing approach to non-destructive depth profiling of deeply buried interfaces by developing a generic and semi-automated method analyzing the inelastic background of XPS spectra



2000: « The interface is the device »
Herbert Kroemer, Nobel Lecture



XPS

- Non-destructive core-level analysis
- Minimal sample preparation
- Limited probing depth

Inelastic Background Analysis (IBA)

- QUASES software package
- 10 nm probing depth with traditional XPS
- no additional instrumentation needed
- Limitations : peak interference in background


HAXPES + IBA

- Up to 150 nm probing depth when combined with background analysis
- Expand collection of available peaks

Objectives

1: Apply IBA to lab-based HAXPES spectra for the characterization of interfaces and buried layers

2: Develop a generalized, semi-automatic method for evaluating results and fitting uncertainty in the IBA solution



Final Objective:

Bring offline HAXPES analysis to inline metrology for semiconductor technologies

Objective 1(a) $\text{Al}_2\text{O}_3 / \text{HfO}_2$



3 bilayer samples of High-k metal oxides with $d(\text{HfO}_2)=2.5 \text{ nm}$

□ Compare Cr K α and Ga K α HAXPES results

Surface layers	Name	Surface Layer Thickness (nm)
Al_2O_3	A1	9.9
Al_2O_3	A2	14.8
Al_2O_3	A3	24.4

- Cr K α $h\nu = 5414.72 \text{ eV}$
IMFP Hf3d = 6.6 - 6.7 nm
- Ga K α $h\nu = 9251.74 \text{ eV}$
IMFP Hf3d = 11.8 - 12.2 nm

Comparison of Interface Positions




- Hf3d is not in accordance with Al1s interface position and Ga K α HAXPES energy did not improve IBA model
- Increasing source energy increases inelastic losses in the kinetic energy range further from the peak (Hf3d), which is interrupted by other peaks (Si1s) in these spectra
- Must carefully choose source energies for individual sample (peak overlap and buried layer thickness)

Objective 1(b) $\text{Al}_2\text{O}_3 / \text{GaN}$


5 bilayer samples representative of HEMT device

□ Apply QUASES-Generate reference spectra




- Start Depth of buried Ga2p too close to surface
- Creates broad interface with Al_2O_3

QUASES-Analyze – no reference



QUASES-Generate – GaN bulk reference spectrum




$$MSE = \frac{1}{n} \sum_{i=1}^n (Y_i - \hat{Y}_i)^2$$


Objective 2. Error Calculation for Evaluation of IBA Solution

Calculate the error between a range of points on the x-axis, and look for the lowest error in possible solutions

Problem: The size and location of the background region for comparison changes with:




Photon Energy




□ Restrict energy range


- Hf3d is not in agreement with Al1s interface position and Ga K α HAXPES energy did not improve IBA model
- HAXPES-IBA solutions from visual inspection and error calculation agree




Buried depth of layer



□ Restrict possible solution



- HAXPES-IBA solutions from visual inspection and error calculation agree
- Visual inspection overestimates start depth



More elegant algorithms are needed for a general error calculation but it could be a useful tool for avoiding user bias in IBA

[1] S. Tougaard, Practical guide to the use of backgrounds in quantitative XPS, J. Vac. Sci. Technol. A. 39 (2021) 011201.

[2] T. R. Bure, O. Renault, E. Nolot, T. Lardin, C. Robert-Goumet, N. Pauly. Assessing advanced methods in XPS and HAXPES for determining the thicknesses of high-k oxide materials: from ultra-thin layers to deeply buried interfaces. Appl. Surf. Sci. 609 (2023) 155317.

[3] O. Renault, P. Deleuze, J. Courtin, R. T. Bure, N. Gauthier, E. Nolot, C. Robert-Goumet, N. Pauly, E. Martinez, K. Artyushkova. New directions in the analysis of buried interfaces for device technology by hard X-ray photoemission. Faraday Discuss., 236 (2022) 288.